

Infrared Thermometer



- IR Temperature: 32°C 42°C / 89°F 108°F
- Accuracy +/- 0.4°C
- Resolution 0.1°C
- Spectral Response 8-14 um
- Measurement Distance 5-15 cm
- Ambient Operating 15°C-35°C
- Response Time approx 1s
- Relative Humidity 10%-90%
- LBD Low Battery Display
- Power Supply 2 x 1.5AAA Battery
- Product Size 148mm x 95mm x 50mm
- Certificates CE, ROHS, ETL, FCC

CE

CODE	DESCRIPTION	PACKAGE
TERINFR	Infrared Thermometer	40 pcs/pack

DIRECTIVE

European Union's Directive 2011/65/EU and EU 2015/863

STANDARDS

EN61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements (Endorsed by AENOR in March of 2013.)

EN61326-2-3:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-3: Particular requirements - Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning (Endorsed by AENOR in March of 2013.)